



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/023,167	12/18/2001	Johannes Catharina Antonius Op De Beek	NL000756	5113
24737	7590	09/09/2005	EXAMINER	
PHILIPS INTELLECTUAL PROPERTY & STANDARDS P.O. BOX 3001 BRIARCLIFF MANOR, NY 10510			STREGE, JOHN B	
			ART UNIT	PAPER NUMBER
			2625	

DATE MAILED: 09/09/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/023,167

Applicant(s)

OP DE BEEK ET AL.

Examiner

John B. Strege

Art Unit

2625

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 21 June 2005.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-11 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 8 and 11 is/are allowed.
- 6) ☒ Claim(s) 1-7,9 and 10 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 18 December 2001 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____.

Response to Amendment

1. The amendment received 6/21/05 has been entered in full. Due to the amendments to the specification the objections regarding the specification have been withdrawn.

Response to Arguments

2. Applicant's arguments filed 6/21/05 have been fully considered but they are not persuasive. Specifically Applicant argues that Cline discloses points and normal vectors thus does not disclose selectively suppressing a geometrically selected part of the data set and rendering an image based on any non-suppressed part of the data set. The Examiner draws the Applicant's attention to col. 2 lines 40-45 which disclose providing a three-dimensional display system for a graphic workstation capable of removing a predetermined volume from a model of an object, and displaying the remaining portions of the model. Thus Cline discloses selectively suppressing a geometrically selected part of the data set and rendering an image based on any non-suppressed part of the data set.

The Applicant further argues that Cline does not teach or disclose that the selected part comprises a first selection containing all points associated to a nearer region with respect to a first clipping plane and moreover all points associated to a farther region with respect to a second clipping plane. Applicant's attention is drawn to col. 4 lines 31-49 where a set of model cut planes 206 are disclosed that comprise a clip box 208 and the user selects a region inside of the clip box 208 to display. Thus Cline teaches a first selection containing all points associated to a nearer region with respect to a first

clipping plane and moreover all points associated to a farther region with respect to a second clipping plane.

Claim Rejections - 35 USC § 102

3. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

4. Claim 1-6 and 9 are rejected under 35 U.S.C. 102(b) as being anticipated by Cline et al. European Patent Application 0549182A2 (as cited in the IDS, hereinafter "Cline").

Cline discloses a method for visualizing a limited part of a 3D medical image-point-related data set, said method being based on selectively suppressing a geometrically selected part of the data set and rendering an image based on any non-suppressed part of the data set (col. 2 lines 41-45 disclose a graphic workstation to remove a predetermined volume from a model of an object and col. 4 lines 8-14 recite that the model may be volume data obtained from a system for providing three-dimensional imaging such as tomography, and col. 2 lines 32-34 recite that the invention may be used for surgery procedures), said method being characterized in that said selected part comprises a first selection containing all points associated to a nearer region with respect to a first clipping plane (set of model cut planes 206, seen in figure 2) and moreover all points associated to a farther region with respect to a second

Art Unit: 2625

clipping plane (set of model cut planes 206, seen in figure 2), respectively, thereby making the rendered image being based on an intermediate region between said first clipping plane and said second clipping plane (col. 4 lines 29-49).

Regarding claim 2, the cutting planes 206 are parallel planes.

Regarding claim 3, as seen in figure 4, the cutting planes are substantially parallel to the imaging plane.

Regarding claim 4, the intermediate region of the cutting planes encompass a region of interest (col. 6 lines 34-35).

Regarding claim 5, Cline discloses user manipulation of the immediate region through rotation of the model (col. 4 lines 39-49).

Regarding claim 6, the cutting planes disclosed by Cline have a diverging angle less than 25 degrees.

Claim 9 is similar to claim 1, except claim 9 is an apparatus claim. As Cline discloses both a method and apparatus the same argument used above for claim 1 applies equally to claim 9.

Claim Rejections - 35 USC § 103

5. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

6. Claims 7 and 10 are rejected under 35 U.S.C. 103(a) as being unpatentable over Cline, European Patent Application 0549182A2 in view of Stiel et al. *Digital Flasing Tomosynthesis: A Promising Technique for Angiocardigraphic Screening* (as cited in the IDS, hereinafter "Stiel").

Claim 7 is dependent on claim 1 (rejected above over Cline) and includes the further limitation – wherein said first and second limited parts are substantially coinciding but associated to respective stereoscopic lines of view, and said method furthermore providing for viewing a 3-d stereo image of said intermediate region through combined viewing of said first and second limited parts. Cline does not explicitly disclose respective stereoscopic viewing of selected parts of the model.

Stiel discloses a tomosynthesis system for stereoscopic viewing of tomoangiograms of single slices or multiple slices of varying thickness (from the abstract and section F starting on page 318).

Cline and Stiel are analogous art because they are from the same field of endeavor of viewing tomographic images.

At the time of the invention it would have been obvious to one of ordinary skill in the art to combine Cline with Stiel to obtain stereoscopic imaging of selective parts of the tomoangiograms. The motivation for doing this would be to eliminate unnecessary data (as disclosed by Cline col. 2 lines 32-34) and allow clearer stereoscopic viewing of the arteries. Thus it would have been obvious to one of ordinary skill in the art to combine Cline and Stiel to obtain the invention as specified in claim 7.

Claim 10 is similar to claim 7, except claim 10 is an apparatus claim. Thus as Cline discloses both a method and apparatus the same argument used above applies equally to claim 10.

Allowable Subject Matter

7. Claims 8 and 11 are allowed. The reasons for allowance were given in the previous office action.

Conclusion

8. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Contact Information

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John B. Strege whose telephone number is (571) 272-


Art Unit: 2625

7457. The examiner can normally be reached on Monday-Friday between the hours of 8-5.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Bhavesh Mehta can be reached on (571) 272-7453. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

JS


BHAVESH M. MEHTA
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 2600